

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

SEP 22 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>AW Anderson</u>	ID# <u>2467</u> Date <u>9/18/09</u>
Agency <u>Seldovia Police Dept.</u>	Phone # <u>234-7640</u>
Instrument Location <u>Seldovia Police Dept</u>	
A	
Alco S/N <u>X301530</u> ✓	Target Value <u>.078</u> High Pressure <u>500 PSI</u>
B	
Alco Test Value Average <u>.075</u>	<u>.075</u>
1 st Alco	2 nd Alco
Signature <u>AW Anderson</u>	
(OVER)	

(Do Not write in the area below)

BMB
10/18/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of Oct, 2009.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



*of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Datamaster cdm S/N 130180 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

SEPTEMBER 18, 2009

OPERATOR'S NAME:

ANDERSON AW

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY VERIFICATION

SUBJECT'S FIRST NAME/MI :

N/A

O.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .075 ✓

ALCO S/N: X301538 ✓

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.65 in	.077	10:21
ALCO TARGET	.000	10:22
BLANK TEST	VERIFIED	10:22
INTERNAL STANDARD	.075	10:22
ALCO TV 29.65 in	.000	10:23
BLANK TEST	.000	10:24
SUBJECT SAMPLE	.000	10:24
BLANK TEST	.075	10:25
ALCO TV 29.65 in	.000	10:25
BLANK TEST		

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

SEPTEMBER 18, 2009
TIME 10:18

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/28/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.65 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~